

MAHARAJA AGRASEN MEDICAL COLLEGE AGROHA (HISAR)

CORRIGENDUM

(Date & Specification - Laparoscopy Machine)

E-Tender is invited vide tender ID No. **2024_HRY_366563_1** for Supply and Installation of Laparoscopy Machine at Maharaja Agrasen Medical College, Agroha (Hisar). The last date of submission is extended up to **08/05/2024** at 17:00 PM. and technical bid will be opened on **09/05/2024** at 11:00AM. Details may be seen at <https://etenders.hry.nic.in/nicgep/app>


Medical Superintendent
For Director

29/4/24

The following amendment is made in specification for Laparoscopy Machine is as under:-

- 1 Under the heading 4K camera System:
Color Gamut BT2020 should be read as Color Gamut BT2020/System Specific
- 2 Under the heading Digital CO₂ Insufflator with heating, there is a subheading. Specifications of Insufflator system
Points g., h., i., j. are a misprint in original Tender and are not relevant so kindly omit.
- 3 Under the heading Endoscopy Trolley
Cart should be either from the same manufacturer as endovision or should be USFDA approved

