

CORRIGENDUM

(Amendment specifications - Physiotherapy Equipment)

An E-Tender is invited for procurement of physiotherapy equipment at Maharaja Agrasen Medical College Agroha (Hisar). The following amendments/modifications is made in the DNIT. Details may be seen at <https://etenders.hry.nic.in/nicgep/app>

Tender ID:- 2024_HRY_367459_1


Medical Superintendent
For Director

Amendment for Specifications of Shock Wave Therapy Device :-

Sub:- Revised Specification for Shock Wave Therapy Device

Focused Shockwave Therapy –


A whole new formula for Chronic Pain

The device must be useful for Orthopedic pain conditions like Frozen Shoulder, Calcaneal Spur, OA Knee, etc. using electro-acoustic sound waves with proven clinical results

General Characteristics:

1. Must be operating with Single Piezoelectric Crystal and an Electroacoustic Lens
2. Highly-effective, non-invasive treatment for pain associated with musculoskeletal system
3. Built-in Therapeutic Encyclopedia - description and images explaining application of all pre-programmed treatment programs
4. Should have pre-programmed protocols
5. Self-identifications of accessories and accessories maintenance check
6. Ergonomic Applicator must eliminate backward shocks
7. Applicator must give at least 3 million shocks
8. Adjustable depth of penetration up to 65 mm with 3 Coupling pads sizes 0-35 mm, 15-50 mm and 30-65 mm
9. Hand piece / Applicator should have display of number of shocks with intensity
10. Must have Touch Screen display on applicator hand piece
11. Should be supplied with Trolley with Castors and Drawers
12. Quality – International Standards / European CE

dep



Nidhi Aggarwal
Physiotherapist