

MAHARAJA AGRASEN MEDICAL COLLEGE AGROHA (HISAR)

CORRIGENDUM (Amendment in Specifications) (Equipments Anesthesia -4)

In continuation of the E-Tender published vide E-Tender ID **2026_HRY_538090_1** for procurement of **equipments** at Maharaja Agrasen Medical College, Agroha (Hisar), it is hereby informed that the specification of the equipment(s) mentioned in the tender has been revised/amended.

Accordingly, the **revised specification** shall be read and considered in place of the specification originally uploaded in the tender document.

Bidders are requested to take note of the revised specification and submit their bids accordingly.

All other terms and conditions of the tender shall remain **unchanged**.

The Revised/Amendment specifications are as under:-


Medical Superintendent
For Director

05-4

Specification for Compendium - Tenda Arrestalis-4

**Technical specifications of portable flexible intubating video bronchoscope/
fiberoptic bronchoscope**

1. Should have minimum insertion cord length 600 mm.
2. Should have depth of field 3-50 mm.
3. Should have field of view minimum 85 degrees.
4. Insertion tube should have around 5 mm outer diameter for adults.
5. Should have bending angles of at least 180 degrees up and 160 degrees down.
6. Should have distal end diameter less than 6.2 mm.
7. Should have minimum instrument channel width of 2.0 mm for adults.
8. Should have high intensity illuminating system.
9. Should be supplied with standard accessories.
10. Should have large high-resolution display.
11. Should have a facility to attach suction tubing and a switch to activate suction through distal tip.
12. Monitor should have facility to attach videolaryngoscope also.

Pranav
Sheetal

Sp
01/11/2016

Tyoti

Vishu

Sumit

Rish
Rohit